Intermountain® Healthcare

HOMECARE SERIES

Wallaby 3 Biliblanket Care and Use Guidelines

For Your Baby's Safety

Following these safety guidelines will help reduce potential risks. Refer to the manufacturer website if you would like further information about your equipment.

- Use the biliblanket only for its intended use.
- Follow your physician's advise when using oxygen with phototherapy.
- Do not operate the biliblanket without the T-Vest (D).
- Do not spill liquid onto the equipment. Electric shock hazard and or equipment failure could result.
- Do not leave your baby unattended during phototherapy.
- Allow air to circulate around the unit.
- Turn the Illuminator Unit (A) off when the baby is not in the T-Vest (D).

When setting up the biliblanket, choose a place with the following features:

- Room temperature at 72°F or higher.
- A sturdy, flat surface away from drafts (bed, crib, or dining table).
- An accessible electrical outlet.
- Easy access for continual caregiver attendance.
- Absence of flammable substance or cleaning agents.

Setting up Your Biliblanket

Equipment Contents



(A) Illuminator Unit
(B) Ferrule
(C) Power cord
(D) T-Vest
(E) Panel

1 Place the **Illuminator Unit (A)** on a hard, flat surface.

2 Insert the **Ferrule (B)** — with the metal peg facing up — into the opening of the **Illuminator Unit (A)**. Twist the **Ferrule (B)** one-quarter turn to lock into place.

3 Plug one end of the **Power Cord (C)** into the back of the **Illuminator Unit (A),** and the other end into an electrical outlet.

4 Flip the **ON/OFF** switch to verify the Panel (E) illuminates; then switch the power off.



5 Slide **Panel (E)** with the sticker side down all the way into the **T-Vest (D)** pocket.



6 Lay the covered **Panel (E)** on a flat surface making sure the sticker side of the panel is facing down.

Placing your baby in the biliblanket

1 Lay the baby on top of the **T-Vest (D)** covered **Panel (E)**. The baby's back should be on the **Panel (E)** with the cable laying between the baby's legs.

2 Wrap the **T-Vest (D)** around the baby and secure with tab.

T-Vest Tips

- Make sure the T-Vest (D) is not wrapped too tight. Your finger should easily fit between the T-Vest (D) and the baby's body.
- Wrap your baby in a blanket or put the baby in a sleeper for a tighter fit.
- For a larger or more active baby, tape the T-Vest (D) to the baby's diaper.
- If the T-Vest (D) gets soiled, discard it and use a new one.

Starting Therapy

1. Ensure the light level on the Illuminator Unit (A) is set at 1 unless instructed different by the doctor.
2. Flip the **ON/OFF** power switch to on.
3. Verify the Panel (E) is lighting under the baby.

Call Intermountain Homecare & Hospice if…

- You experience any problems with the equipment.

Troubleshooting

If you have questions about the equipment or if the equipment needs repair, call Intermountain Homecare & Hospice at 1-800-527-1118. For additional information, please see the user's manual.

Problem	Action
Power light is not green.	• Make sure power cord is properly attached and plugged into an active electrical outlet • Verify the unit is turned on.
System Alert LED is blinking yellow	• One light bulb has burned out and the system has automatically switched to the second light bulb. Continue therapy.
System Alert LED is solid yellow	• Unit may have overheated or both light bulbs have burned out. Allow unit to cool. If condition persists call Intermountain Homecare & Hospice.
Light is not emitting from the fiber optic panel.	• Make sure the panel is securely locked into illuminator box. • Check System Alert LED. If solid yellow, both lights are burned out or unit has overheated. Allow unit to cool. If condition persists, call Intermountain Homecare & Hospice.

Intermountain® Homecare & Hospice

Content created and approved by Intermountain Homecare & Hospice

© 2012–2014 Intermountain Healthcare. All rights reserved. The content presented here is for your information only. It is not a substitute for professional medical advice, and it should not be used to diagnose or treat a health problem or disease. Please consult your healthcare provider if you have any questions or concerns. More health information is available at **intermountainhealthcare.org. Patient and Provider Publications 801-442-2963 FSHC006 - 09/14**.