

Giving an Enteral Feeding with a Joey Pump (Feed Only)

Getting ready

- 1 Fill the feeding bag with formula before setting up the pump.
- **2 Turn the pump on** by pressing the power button on the lower right-hand corner.
- 3 Select:
 - Keep settings for day-to-day use with the same feeding routine

OR

- Clear settings to enter new pump settings
- 4 Load the "Feed Set" according to the manufacturer's diagram and instructions. Don't overstretch the tubing.



- 5 Close the door. The pump will read SET LOADED if the tube is loaded correctly.
- 6 If using an IV pole, position the fluid line in the feeding set bag so that it is 6 inches above the top of the feeding pump.

Replace the pump sets (bag and tubing) every 24 hours. Clean the pump regularly with a damp cloth. Dry well.

Priming the pump

- 1 Disconnect the pump from the patient.
- 2 Press **Prime Pump** to access the pump priming options:
 - Automatic priming of tubing:
 Press Auto Prime. The pump primes the line and stops before the formula gets to the end of the feeding line.
 - Manual control priming of tubing: Press Hold to prime feed until the tubing is primed with formula.
 - Press Done.

Set the feed rate:

- 1 Select Adjust Feed and then Feed Rate.
- 2 Use the buttons on the left of the screen to program the feed rate in amounts ranging from 1 to 400 mL.
- 3 Select Enter when the desired rate is set.
- 4 Select Feed VTBD (volume to be delivered).
- 5 Use the buttons on the left of the screen to program the VTBD to 0 (zero) unless told otherwise by your doctor.
- 6 Select Enter and then Done.

Scan the code or visit the link below to watch a video.

https://app.shoreline. health/s/VLYbCFGTkyl9




Flush the feeding tube

Follow the amounts listed on the tube handout.

Connect the pump set to your feeding tube

- Remove the protective cover from the end of the bag tubing.
- Use the white cone adapter or remove it and use the ENFit connection to connect to your tube.



To start the feeding:

Press Run to start the feeding. A small droplet symbol (•) will appear and will scroll or run down the screen while feeding.

To clear the volume or change the rate after the pump has started:

- 1 Select Hold.
- 2 Select Clear Volume to clear the volume of the amount that has been fed.
- 3 Select Adjust Settings to adjust all settings.
- 4 Select Run to return to normal operations.

When the bag is empty and the pump alarms as a feed error:

- 1 Press Stop Alarm.
- 2 Disconnect the feeding line from the patient.
- 3 Rinse out the bag with warm water.
- 4 Refill the bag with fresh formula.
- 5 Press Continue to start the pump running.

Battery operation

Unplugging the pump from the AC line will automatically place the pump on battery power.

Recharging the battery

- To recharge the battery, plug the pump into an AC wall outlet. The pump will begin to charge.
- When attaching the a/c adapter power cord to the pump, make sure there is a "click" sound The battery symbol should be solid black when fully charged ().

To look at the history, change the alarm loudness, or change the language do the following:

- 1 Press the More button.
- 2 Select **Buzzer** to increase or decrease the volume.
- 3 Select **History** to see how much has been delivered.
- **4** Select **Language** to change the language.



Troubleshooting guide for alarms

Hold Error	 The pump is in holding mode for more than 10 minutes. Press Continue to return to the holding screen, or press and hold the power button to turn off the pump. From the holding screen, you can also Run or Adjust Settings.
Flow Error	 Formula is not being delivered because of a clog or block between the pump and the patient (this is called downstream occlusion). Check for and correct the blockage. Replace the bag or tubing if needed.
Feed Error	 Feeding is not being delivered because the bag is empty or there is a clog between the pump and the feeding bag (upstream occlusion). Check the delivery set tubing to see if it is pinched or kinked, and then reposition. If the occlusion cannot be cleared, load a new bag or tubing set, prime it, and press Continue to restart the feeding.
Flow Error Use > 24 Hours	 Pump set in use beyond the recommended time. An informational warning message blinks on the upper left of the screen. It is recommended to replace the bag every 24 hours.
Pump Set Dislodged	 Pump set is not properly loaded. Check and correct the positioning of the tubing around the rotor. Load a new bag or tubing set if the error cannot be resolved.
Battery Low	Sounds when there is approximately 15 minutes of battery life remaining indicating the battery must be recharged immediately. Plug into AC power to recharge.
Rotor Error	 A problem with the pump set tubing around the pump rotor. Reload the tubing so that it is placed in the middle of the rotor and secure.
System Error	Power down and restart the pump to clear the error. Call the Enteral Team if the error does not clear.

Need help or have questions?

Enteral Team

(385) 887-7356

Intermountain Homecare Main Numbers

(385) 887-6000 | (800) 527-1118 (Toll Free)

Content created and approved by Intermountain Homecare & Hospice